

Column Care and Use Instructions

YMC-Pack Diol

Thank you for purchasing this product. To ensure optimal performance and durability of the column, please read these instructions carefully before using this column.

1. Specifications

Column	Particle size (μm)	Pore size (nm)	Functional group	Usable pH range	Temperature (Upper limit)
Diol-60	3, 5	6	Dihydroxypropyl	5–7.5	40°C
Diol-120		12			
Diol-200	2, 3, 5	20			
Diol-300		30			

2. Shipping solvent

The column is shipped in the 0.05% sodium azide aqueous solution. Flush with water sufficiently before replacing with the mobile phase.

3. Precautions for use

- The correct direction of the solvent flow is indicated by an arrow on the column identification label.
- Tubing must have flat ends and must bottom out in the column endfitting. Tubing must be connected to the column correctly to avoid creating a void between the column frit and tubing, which can cause a leak and result in poor column performance (e.g. peak tailing, loss of theoretical plate number).
- The column pressure limit and recommended flow rate are the following.

Particle size	Pressure limit
2 μm	45 MPa (30 MPa or less for regular use)
3 μm, 5 μm	20 MPa

- ※ Avoid using a column repeatedly near the pressure limit or abrupt change in pressure to prevent shortening of the column life.
- ※ Adjust the flow rate appropriately because the pressure changes depending on the column length, temperature, types of mobile phase etc.
- Aqueous mobile phase is basically used. Total salt concentration of mobile phase should be lesser than 0.7 M. Tris-HCl, citrate, etc. are applicable as buffer solution. These buffer solutions are available with solutions containing buffer salt/additives such as sodium chloride, sodium sulfate, and ammonium sulfate.
- Aqueous solutions of urea and guanidine hydrochloride which are used for a denaturant of the protein can be used. Moreover, 0.1% or less concentration of surface-active agents such as Tween80, SDS is also usable. When using these mobile phases, the equilibration of the column needs long time as compared to the general mobile phase.
- Methanol or acetonitrile can be added to mobile phase. When using a mobile phase containing methanol or acetonitrile, be mindful of the operation pressure rising by increasing viscosity and the precipitation of buffer salts/additives.
- Recommendations of pH and temperature for column use are shown in the specifications table in section 1.
- Column lifetime varies depending on conditions of use such as pH, temperature and mobile phase composition. In general, usage at higher temperatures and higher concentrations of buffer salts/additives can shorten the column lifetime.
- For storage except daily use, the column should be flushed thoroughly with water, replaced in aqueous solution of 0.05% of sodium azide, which should be sealed the both ends tightly and stored in a location with minimal temperature change. In next time, flush with water sufficiency before replacing with the mobile phase.

4. Column cleaning (general method)

In the case that some hydrophobic proteins or the hydrophobic materials are adsorbed or retained, flush the column with solvent containing high salt concentration (approx. 0.5 M). At that time, be careful about usable pH.

5. Other environments

- The operating pressure should be kept under 5 MPa (725 psi) for Glass column.
- To prevent exposure of the column to excessive pressure, the sample solution should be filtered through a 0.2 μ m membrane or smaller to remove particulates. We recommend using a pre-column filter to prevent the column frit from being clogged with samples.

● Please note that specifications and appearance may change without notice.